

NANOCOLOR[®] gesamt-Phosphat 15

total Phosphate / Phosphate total / Fosfato total



Test 0-80
REF 985080

Methode / Method / Méthode / Método

0801 0.30–15.00 mg/L P

0802 1.0–45.0 mg/L PO₄³⁻

0803 0.7–34.5 mg/L P₂O₅

690 / 710 nm

Min

20–80 %

Max

0.30











3.24 12.06

15.00 mg/L P

1.0

9.8 36.2

45.0 mg/L PO₄³⁻



NANOCOLOR[®] ortho-Phosphat 15

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-80
REF 985080

Methode / Method / Méthode / Método
 0805 0.30–15.00 mg/L PO₄-P
 0806 1.0–45.0 mg/L PO₄³⁻
 0804 0.7–34.5 mg/L P₂O₅

690 / 710 nm

Min	20–80 %		Max
0.30	3.24	12.06	15.00 mg/L PO ₄ -P
1.0	9.8	36.2	45.0 mg/L PO ₄ ³⁻

